

CLINICAL EVALUATION OF THYROXINE RADIOIMMUNOASSAY
BASED ON THE SOLID PHASE TECHNIQUE

Minoru Fukuchi, Akira Kido, Horoaki Kitani, Akiharu
Nishikawa, Shuitsu Satake, Toshitaka Morita and
Kiyoyasu Nagai

RI Center, Hyogo College of Medicine, Nishinomiya.

Rapid method of the determination of serum thyroxine was developed recently based on the solid phase technique for preparation of bound(B) and free(F). This paper presents the results of clinical evaluation on the solid phase radioimmunoassay system for thyroxine using SPAC T₄ and GammaCoat T₄ kits in 99 normal subjects and 201 patients with various thyroid diseases.

Multiple dilutions of patients sera resulted in curves parallel to that obtained by standard T₄ in both kits. The intraassay reproducibility(C.V.) of serum T₄ assays was 6.54% in SPAC T₄ kit, and 2.05% in GammaCoat T₄ kit. The interassay reproducibility (C.V.) of serum T₄ assays was 6.40% in SPAC T₄ kit, and 4.05% in GammaCoat T₄ kit. The B % of standard curves and serum samples was increased significantly correlated to the incubation times. Good correlation was found between assay results of SPAC T₄ kit and GammaCoat T₄ kit - the correlation coefficient was $r = +0.94$, $y = 1.05x + 0.48$.

Serum T₄ levels of normal subjects was ranged from 4.6 to 12.1 µg/dl with SPAC T₄ kit, and from 5.3 to 11.3 µg/dl with GammaCoat T₄ kit. The assay results of hyperthyroid sera and hypothyroid sera was significantly different from euthyroid sera in both kits.

The accuracy and specificity of these systems is thought to be very useful for routine clinical determinations of thyroxine.

EVALUATION AND CLINICAL USE OF SOLID PHASE
RADIOIMMUNOASSAY FOR THYROXINE

Keiko Nishi*, Kiyoshi Miyai* and Minoru Kawashima**

*Central Laboratory for Clinical Investigation, Osaka University Medical School, and ** Osaka Kessei Laboratories, Osaka.

Two types of solid phase radioimmunoassay of thyroxine (T₄) using Gammacoat T₄ (G) and Spac T₄ (S) Kits were evaluated for clinical application, and the following results were obtained, (1) Standard curves were compared when incubation was carried out at various temperatures (4, 26, 37°C) for 45 min. B/T ratio increased at the low level of T₄ when the incubation temperature was elevated, but T₄ values in three kinds of pooled serums determined at the same temperature were constant. (2) When the incubation time was varied from 15 min to 3 hours at constant room temperature, again, B/T ratio increased at the low level but serum value of T₄ determined at the same condition was constant. (3) Dilution of serum resulted in the parallel curve with that of standard T₄. (4) When ¹²⁵I-T₄ was reduced to 2/5, B/T ratio was increased but values of serum T₄ determined at the same condition were constant, (5) The recoveries of standard T₄ added to the pooled serum were 106 ± 3.3% (G) and 100 ± 2.9% (S) respectively. (6) Coefficients of variation within assay were 3.2 - 4.7% (G), 2.4 - 5.6% (S) and those between assay were 2.6 - 4.1% (G) and 3.4 - 6.5% (S) respectively. (7) Serum with slight hemolysis had no significant effect on T₄ value. (8) Serum concentrations of T₄ determined by means of these methods were 8.1 ± 1.57 µg/dl in 69 normal subjects (G) and increased in hyperthyroid and decreased in hypothyroid patients. (9) T₄ values determined by the double antibody method of Eiken (X) was well correlated with those determined by the solid phase methods (Y, Y) as follows: $Y = 0.99X + 0.94$ ($r=0.98$) $Y = 1.05X + 0.17$ ($r=0.97$), In conclusion, the solid phase radioimmunoassay of T₄ was useful for clinical test.