

Improved image quality with elliptical orbits and distance-weighted backprojection SPECT reconstruction

R. L. EISNER, Ph.D.,* W. A. FAJMAN, M.D.,* D. J. NOWAK, M.S.**
and R. I. PETTIGREW, M.D., Ph.D.*

**Emory University School of Medicine and Grady Memorial Hospital, Atlanta, Georgia*

***General Electric Company, Medical Systems Group, Milwaukee, Wisconsin*

The image quality of SPECT bone scans is enhanced through the use of distance-weighted filtered backprojection in association with elliptical orbit acquisition.

Key words: Radionuclide studies, Single photon emission computed tomography, Reconstruction techniques